



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Canyon Labs Holdings, LLC
2865 Scott St STE 103
Vista, CA 92081

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at www.anab.org.



Jason Stine, Vice President

Expiry Date: 20 January 2028

Certificate Number: AT-3455



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Canyon Labs Holdings, LLC

2865 Scott St STE 103

Vista, CA 92081

Steve Greene sgreene@canyonlabs.com
760 505 2724

TESTING

ISO/IEC 17025 Accreditation Granted: **20 January 2026**

Certificate Number: **AT-3455**

Certificate Expiry Date: **20 January 2028**

Mechanical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Seal Strength for Flexible Barrier Materials	ASTM F88	Packaging Materials, Packages and Components	Steel Ruler, Tensile Machine
Tensile Properties of Thin Plastic Sheeting	ASTM D882		Thickness Gage, Steel Ruler, Tensile Machine
Pressure Decay Leak Test for Flexible Packages with or without Restraining Plates	ASTM F2095 (Method A)		Pressure Delivery System with Timer Controls
Accelerated Aging of Sterile Barrier Systems for Medical Devices	ASTM F1980		Temperature and Humidity Conditioning Chamber
Internal Pressurization Failure Resistance of Unrestrained Packages	ASTM F1140 (except 9.1.1)		Pressure Delivery System with Timer Controls
Detecting Seal Leaks in Porous Medical Packaging by Dye Penetration	ASTM F1929		Dye Dispenser, Dye Solution Container, Cutting Instrument, Eyedropper, Aqueous Dye Penetrant Solution.
Detecting Gross Leaks in Packaging by Internal Pressurization (Bubble Leak)	ASTM F2096		Pressure Delivery System
Conditioning and Testing Flexible Barrier Materials	ASTM E171		Temperature and Humidity Conditioning Chamber
Thickness Measurement of Flexible Packaging Materials	ASTM F2251		Thickness Gage

This Scope of Accreditation, version 001, was last updated on: 20 January 2026 and is valid only when accompanied by the Certificate.

Page 1 of 2

1899 L Street NW, Suite 1100-A, Washington, DC 20036

414-501-5494

www.anab.org

Mechanical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Conditioning Containers, Packaging or Packaging Components for Testing	ASTM D4332	Packaging Materials, Packages and Components	Temperature and Humidity Conditioning Chamber
Detecting Leaks in Nonporous Packaging or Flexible Barrier Material by Dye Penetration	ASTM F3039		Dye Dispenser, Absorbent Surface, Aqueous Dye Penetrant Solution.
Determining Integrity of Seals for Flexible Packaging by Visual Inspection	ASTM F1886		Light Source, Light Meter
Abrasion Resistance of Printed Materials by the Sutherland Rub Tester	ASTM D5264		Sutherland Rub Tester
Slow Rate Penetration Resistance of Flexible Barrier Films and Laminates	ASTM F1306		Caliper, Thickness Gage, Penetration Probe, Tensile Machine
Package Yield of Plastic Film	ASTM D4321		steel rule, caliper, weight scale
Temperature Test for Transport Packaging	ISTA 7D		Temperature Conditioning Chamber
Testing Standard for Thermal Transport Packaging Used in Parcel Delivery System Shipment	ISTA 7E		Temperature Conditioning Chamber



Jason Stine, Vice President